

# HeartMate 3™ Pocket Controller Flashing Yellow Diamond Alarm

---

## What is a Flashing Yellow Diamond Alarm?

- The Yellow Diamond Alarm indicates low external battery power.

## What does the Flashing Yellow Diamond Alarm mean?

- This is an **Advisory Alarm**.
- You have less than 15 minutes of battery power remaining.
- This alarm occurs because the System Controller is not getting enough power or voltage.

## What does the Flashing Yellow Diamond Alarm sound like?

The yellow diamond alarm is a “slow beeping sound”.

## How do I respond to the Flashing Yellow Diamond Alarm?

If you experience a flashing yellow diamond alarm, you will need to connect to a different power source. You have two power source options:

- Change to a new set of batteries

**OR**

- Switch power source to the Mobile Power Unit.

If the alarm continues after you try these steps, call the VAD emergency number 734-477-6722.

**Reminder:**  
Always check the fuel gauge when changing your batteries.

Disclaimer: This document contains information and/or instructional materials developed by Michigan Medicine for the typical patient with your condition. It may include links to online content that was not created by Michigan Medicine and for which Michigan Medicine does not assume responsibility. It does not replace medical advice from your health care provider because your experience may differ from that of the typical patient. Talk to your health care provider if you have any questions about this document, your condition or your treatment plan.

Author: Center for Circulatory Support Team  
Reviewers: Bethany Lee-Lehner, RN, MSN

Patient Education by [Michigan Medicine](#) is licensed under a [Creative Commons Attribution-NonCommercial-ShareAlike 4.0 International Public License](#). Last Revised 02/2017  
CVC Control #1102