

# Heartware™ Changing Your System Controller

---

## How do I change the controller?

1. Have the patient lay down for safety.
2. Prepare the new controller:
  - a. Connect 2 new batteries to the controller.
  - b. Pause for approximately 10 seconds while the new controller powers up and sounds an alarm.
3. Pull back the white driveline cover from the original controller's silver connection.
4. Disconnect the driveline from the original controller by grasping the ridges and pulling out the silver connector.
5. Connect the driveline to the new controller by aligning the two red marks and plugging it in.
6. Insert the red alarm adapter into the blue connector port on the original controller.
7. Disconnect both power sources from the original controller. The controller will turn off and the alarms will stop.
8. Slide the white driveline cover over the new controller's silver connection.

Disclaimer: This document contains information and/or instructional materials developed by Michigan Medicine for the typical patient with your condition. It may include links to online content that was not created by Michigan Medicine and for which Michigan Medicine does not assume responsibility. It does not replace medical advice from your health care provider because your experience may differ from that of the typical patient. Talk to your health care provider if you have any questions about this document, your condition or your treatment plan.

Author: Kelly Lawrence, OTR/L  
Reviewers: Bethany Lee-Lehner, RN, MSN  
Center for Circulatory Support Team

Patient Education by [Michigan Medicine](#) is licensed under a [Creative Commons Attribution-NonCommercial-ShareAlike 3.0 Unported License](#). Last Revised 02/2018  
CVC Control #1108

Center for Circulatory Support & Occupational Therapy  
Heartware™ Changing Your Controller